



Product Service

Attestation of Conformity

No. N8A 089832 0022 Rev. 00

Holder of Attestation: **SUZHOU ANTAI AIRTECH CO., LTD**


No.2 Weixin Road, SIP
215122 Suzhou, Jiangsu
PEOPLE'S REPUBLIC OF CHINA

Product: **Laboratory Equipment**
Biological safety Cabinet

This Attestation of Conformity is issued on a voluntary basis according to the Low Voltage Directive 2014/35/EU relating to electrical equipment designed for use within certain voltage limits. It confirms that the listed equipment complies with the principal protection requirements of the directive and is based on the technical specifications applicable at the time of issuance. It refers only to the particular sample submitted for conformity assessment. For details see: www.tuvsud.com/ps-cert

Test report no.: 4842024057600

Date, 2024-06-11



(Bin Li)

Page 1 of 2

This Attestation does not replace the regulatory EU Declaration of Conformity (DoC) and does not allow for CE marking. After preparation of the necessary documentation and establishing compliance to requirements of all applicable directives, the manufacturer may sign a DoC and apply the CE marking. The DoC is issued under the sole responsibility of the manufacturer.



Product Service

Attestation of Conformity

No. N8A 089832 0022 Rev. 00

Model(s): BSC-1006IIA2, BSC-1306IIA2,
BSC-1606IIA2, BSC-1806IIA2,
BSC-1006IIB2, BSC-1306IIB2,
BSC-1606IIB2, BSC-1806IIB2

Parameters:

Rated voltage	220-240VAC
Rated frequency	50/60Hz
Rated power	1.65KVA (BSC-1006IIA2, BSC-1006IIB2); 1.75KVA (BSC-1306IIA2, BSC-1306IIB2); 1.85KVA (BSC-1606IIA2, BSC-1806IIA2, BSC-1606IIB2, BSC-1806IIB2)
Protection class	I
Degree of protection	IP20

Tested according to: EN 61010-1:2010/A1:2019

Page 2 of 2

This Attestation does not replace the regulatory EU Declaration of Conformity (DoC) and does not allow for CE marking. After preparation of the necessary documentation and establishing compliance to requirements of all applicable directives, the manufacturer may sign a DoC and apply the CE marking. The DoC is issued under the sole responsibility of the manufacturer.