



Product Service

# Attestation of Conformity

No. E8A 123678 0002 Rev. 00

**Holder of Attestation:** **Qingdao Antech Scientific Co., Ltd.**

No.182 Haier Road, Laoshan District  
266000 Qingdao, Shandong  
PEOPLE'S REPUBLIC OF CHINA

**Name of Object:** **Laboratory Equipment  
(Pharmacy Refrigerator)**

This Attestation of Conformity is issued on a voluntary basis according to the Directive 2014/30/EU relating to electromagnetic compatibility. It confirms that the listed apparatus complies with the principal protection requirements of the directive and is based on the technical specifications applicable at the time of issuance. It refers only to the particular sample submitted for conformity assessment. For details see: [www.tuvsud.com/ps-cert](http://www.tuvsud.com/ps-cert)

**Test report no.:** 64711210159601A

**Date,** 2023-12-11

( Tony Liu )

Page 1 of 2

This Attestation does not replace the regulatory EU Declaration of Conformity (DoC) and does not allow for CE marking. After preparation of the necessary documentation and establishing compliance to requirements of all applicable directives, the manufacturer may sign a DoC and apply the CE marking. The DoC is issued under the sole responsibility of the manufacturer.



TUV®



Product Service

# Attestation of Conformity

No. E8A 123678 0002 Rev. 00

**Model(s):** MPR-1200, MPR-525, MPR-60, MPR-60S, MPR-650, MPR-650S, MPR-120, MPR-120S, MPR-280, MPR-280S, MPR-362, MPR-362S, MPR-440, MPR-440S

**Description of Object:**

Rated Voltage: 220-240V~  
Rated Frequency: 50Hz  
Rated current:  
1.2A for MPR-1200, MPR-525, MPR-60, MPR-60S, MPR-650, MPR-650S, MPR-120, MPR-120S  
1.5A for MPR-280, MPR-280S  
1.6A for MPR-362, MPR-362S, MPR-440, MPR-440S  
Protection class: I

**Tested according to:** EN 61326-1:2013

Page 2 of 2

This Attestation does not replace the regulatory EU Declaration of Conformity (DoC) and does not allow for CE marking. After preparation of the necessary documentation and establishing compliance to requirements of all applicable directives, the manufacturer may sign a DoC and apply the CE marking. The DoC is issued under the sole responsibility of the manufacturer.



TUV®